

Display Report

Analysis Info

Analysis Name D:\Data\TEST\1107\manguogan_86_01_330.d
Method 330.m
Sample Name manguogan
Comment

Instrument amaZon SL

Acquisition Parameter

Ion Source Type	ESI	Ion Polarity	Negative	Alternating Ion Polarity	Off
Mass Range Mode	UltraScan	Scan Begin	70 m/z	Scan End	1000 m/z
Capillary Exit	-140.0 V	n/a	n/a	Trap Drive	79.4 V
Accumulation Time	122424 μ s	Averages	5 Spectra	Auto MS/MS	Off

